



## AssayMax Human Apolipoprotein B ELISA Kit

Catalog Number EA7001-1

### Introduction

Apolipoprotein B (Apo B) is the dominant protein constituent of LDL. The levels of secreted apoB directly correlate with circulating serum cholesterol levels (1). Apo B is a better marker of risk of vascular disease than other lipid markers including LDL and HDL-cholesterol and triglycerides (2). Apo B is consistently associated with an increased mortality in type 1 diabetes (3). Plasma apo B and VLDL and LDL with apo B are independent risk factors for cardiovascular disease (CVD) (4). Apo B, Apo A-I and the Apo A-I/Apo B ratio can predict incident ischemic stroke among patients with preexisting atherothrombotic disease (5).

### Principal of the Assay

The AssayMax Human Apo B ELISA (Enzyme-Linked Immunosorbent Assay) kit is designed for detection of human Apo B in plasma, serum, urine, and cell culture samples. This assay employs a quantitative sandwich enzyme immunoassay technique that measures human Apo B in less than 4 hours. A polyclonal antibody specific for human Apo C-II has been pre-coated onto a 96-well microplate with removable strips. Apo B in standards and samples is sandwiched by the immobilized antibody and biotinylated polyclonal antibody specific for Apo B, which is recognized by a streptavidin-peroxidase conjugate. All unbound material is then washed away and a peroxidase enzyme substrate is added. The color development is stopped and the intensity of the color is measured.

### Caution and Warning

- This kit is for research use only.
- The kit should not be used beyond the expiration date.
- The Stop Solution is an acid solution

### Reagents

- **Human Apo B Microplate:** A 96-well polystyrene microplate (12 strips of 8 wells) coated with a polyclonal antibody against human Apo B.
- **Sealing Tapes:** Each kit contains 3 pre-cut, pressure-sensitive sealing tapes that can be cut to fit the format of the individual assay.
- **Human Apo B Standard:** Human Apo B in a buffered protein base (0.4 ug, lyophilized).
- **Biotinylated Apo B Antibody (100x):** A 100-fold concentrated biotinylated polyclonal antibody against Apo B (80 µl).
- **EIA Diluent Concentrate (10x):** A 10-fold concentrated buffered protein base (30 ml).

- **Wash Buffer Concentrate (20x):** A 20-fold concentrated buffered surfactant (30 ml).
- **Streptavidin-Peroxidase Conjugate (SP Conjugate):** A 100-fold concentrate (90 µl).
- **Chromogen Substrate:** A ready-to-use stabilized peroxidase chromogen substrate tetramethylbenzidine (8 ml).
- **Stop Solution:** A 0.5 N hydrochloric acid to stop the chromogen substrate reaction (12 ml).

## Storage Condition

- Store unopened kit at 2-8<sup>0</sup>C up to expiration date.
- Opened EIA Diluent may be stored for up to 1 month at 2-8<sup>0</sup>C. Store reconstituted reagents at -20<sup>0</sup>C or below.
- Opened unused strip wells may return to the foil pouch with the desiccant pack, reseal along zip-seal. May be stored for up to 1 month in a vacuum desiccator.

## Other Supplies Required

- Microplate reader capable of measuring absorbance at 450 nm
- Pipettes (1-20 µl, 20-200 µl, 200-1000µl and multiple channel)
- Deionized or distilled reagent grade water

## Sample Collection, Preparation and Storage

- **Plasma:** Collect plasma using one-tenth volume of 0.1 M sodium citrate as an anticoagulant. Centrifuge samples at 2000 x g for 10 minutes and assay. Dilute samples 1:8000 into EIA Diluent as follows: add 5 µl of sample to 495 µl of EIA Diluent (1:100) to make Solution A; then add 10 µl of Solution A to 790 µl of EIA Diluent (1:80) to make a final working solution (1:8000). Store samples at -20<sup>0</sup>C or below for up to 3 months. Avoid repeated freeze-thaw cycles. (EDTA or Heparin can also be used as anticoagulant.)
- **Serum:** Samples should be collected into a serum separator tube. After clot formation, centrifuge samples at 2000 x g for 10 minutes. Dilute samples 1:8000 into EIA Diluent as follows: add 5 µl of sample to 495 µl of EIA Diluent (1:100) to make Solution A; then add 10 µl of Solution A to 790 µl of EIA Diluent (1:80) to make a final working solution (1:8000). Store samples at -20<sup>0</sup>C or below for up to 3 months. Avoid repeated freeze-thaw cycles.
- **Cell Culture Supernatants:** Centrifuge cell culture media at 2000 x g for 10 minutes to remove debris. Collect supernatants and assay. Store samples at -20<sup>0</sup>C or below. Avoid repeated freeze-thaw cycles.

## Reagent Preparation

- Freshly dilute all reagents and bring all reagents to room temperature before use. If crystals have formed in the concentrate, mix gently until the crystals have completely dissolved.
- **EIA Diluent Concentrate (10x):** Dilute the EIA Diluent 1:10 with reagent grade water. Store for up to 1 month at 2-8<sup>0</sup>C.
- **Standard Curve:** Reconstitute the 0.4 ug of Apo B Standard with 0.8 ml of EIA Diluent to generate a solution of 0.5 ug/ml. Allow the standard to sit for 10 minutes with gentle agitation prior to making dilutions. Prepare triplicate standard points by serially diluting the standard solution (0.5 ug/ml) 1:2 with equal volume of EIA Diluent to produce 0.25, 0.125, 0.063, 0.031 and 0.016 ug/ml solutions. EIA Diluent serves as the zero standard (0 ug/ml). Any remaining solution should be frozen at -20<sup>0</sup>C.

Standard Point	Dilution	[Apo B] (ug/ml)
P1	Standard (0.5 ug/ml)	0.5
P2	1 part P1 + 1 part EIA Diluent	0.25
P3	1 part P2 + 1 part EIA Diluent	0.125
P4	1 part P3 + 1 part EIA Diluent	0.063
P5	1 part P4 + 1 part EIA Diluent	0.031
P6	1 part P5 + 1 part EIA Diluent	0.016
P8	EIA Diluent	0.000

- **Biotin Apo B Antibody (100x):** Spin down the antibody briefly and dilute the desired amount of the antibody 1:100 with EIA Diluent. Any remaining solution should be frozen at -20°C.
- **Wash Buffer Concentrate (20x):** Dilute the Wash Buffer Concentrate 1:20 with reagent grade water.
- **SP Conjugate (100x):** Spin down the SP Conjugate briefly and dilute the desired amount of the conjugate 1:100 with EIA Diluent. Any remaining solution should be frozen at -20°C.

## Assay Procedure

- Prepare all reagents, working standards and samples as instructed. Bring all reagents to room temperature before use. The assay is performed at room temperature (20-30°C).
- Remove excess microplate strips from the plate frame and return them immediately to the foil pouch with desiccant inside. Reseal the pouch securely to minimize exposure to water vapor and store in a vacuum desiccator.
- Add 50 µl of Apo B standard or sample per well. Cover wells with a sealing tape and incubate for two hours. Start the timer after the last sample addition.
- Wash five times with 200 µl of Wash Buffer. Invert the plate and decant the contents, and hit it 4-5 times on absorbent paper towel to completely remove liquid at each step.
- Add 50 µl of Biotinylated Apo B Antibody to each well and incubate for one hour.
- Wash five times with 200 µl of Wash Buffer. Invert the plate and decant the contents, and hit it 4-5 times on absorbent paper towel to complete remove liquid at each step.
- Add 50 µl of Streptavidin-Peroxidase Conjugate to each well and incubate for 30 minutes. Turn on the microplate reader and set up the program in advance.
- Wash five times with 200 µl of Wash Buffer.
- Add 50 µl of Chromogen Substrate per well and incubate for about 10 minutes or till the optimal blue color density develops. Gently tap plate to ensure thorough mixing and break the bubbles in the well with pipette tip.
- Add 50 µl of Stop Solution to each well. The color will change from blue to yellow.
- Read the absorbance on a microplate reader at a wavelength of 450 nm **immediately**. Please note that after the reaction is stopped for about 10 minutes, some black particles may be generated at high concentration point, which will reduce the readings.


## Data Analysis

- Calculate the mean value of the triplicate readings for each standard and sample.
- To generate a Standard Curve, plot the graph using the standard concentrations on the x-axis and the corresponding mean 450 nm absorbance on the y-axis. The best-fit line can be determined by regression analysis using log-log or four-parameter logistic curve-fit.

- Determine the unknown sample concentration from the Standard Curve and multiply the value by the dilution factor.

### Standard Curve

- The curve is provided for illustration only. A standard curve should be generated each time the assay is performed.



### Performance Characteristics

- The minimum detectable dose of Apo B is typically 20 ng/ml.
- Intra-assay and inter-assay coefficients of variation were 4.4% and 7.5% respectively.

## Linearity

Sample Dilution	Average Percentage of Expected Value	
	Plasma	Serum
1:4000	95%	98%
1:8000	105%	103%
1:16000	111%	117%

## Recovery

Standard Added Value	20-200 ng/ml
Recovery %	86 - 110
Average Recovery %	98

## Cross-Reactivity

Species	% Cross Reactivity
Beagle	None
Bovine	None
Monkey	< 10 (suggest 1:100 dilution for plasma)
Mouse	None
Rat	None
Swine	None

## References

- (1) Brodsky JL *et al.* (2008) *Trends Endocrinol Metab.* Sep;19(7):254-9.
- (2) Adiloglu AK *et al.* (2005) *Acta Cardiol.* Dec;60(6):599-604.
- (3) Stettler C *et al.* (2006) *J Intern Med.* Sep;260(3):272-80.
- (4) Furtado JD *et al.* (2008) *Am J Clin Nutr.* Jun;87(6):1623-30.
- (5) Koren-Morag N *et al.* (2008) *J Neurol Sci.* Jul 15;270(1-2):82-7.

Version 2.1